
Midmark M11 - v1.0.5
Total Cycles: 156

Name: RITTER2
Sterilizer ID: V2313228
07/05/2021 06:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.0 F 0.0
2:00 80.1 F 0.2
4:00 90.8 F 0.9
6:00 105.0 F 3.1
8:00 129.6 F 5.8
10:00 155.7 F 8.6
12:00 177.9 F 11.9
14:00 214.6 F 3.7
16:00 216.7 F 7.3
18:00 223.8 F 11.2
20:00 229.1 F 6.9
22:00 243.4 F 12.2
24:00 239.0 F 10.4
26:00 251.4 F 16.4
28:00 262.1 F 22.5
30:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.0 F 29.0
2:00 272.0 F 29.4
3:00 271.8 F 29.7
4:00 271.2 F 28.9

Min 271.2 F 28.8
Max 272.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 229.8 F 0.0
10:00 204.1 F 0.0
15:00 201.2 F 0.0
20:00 206.6 F 0.0
25:00 208.6 F 0.0
30:00 210.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:35
CYCLE COMPLETE

07/05/2021 07:51 AM
ACCEPTED:
REJECTED: X
Reason: Wet Instruments
Operator: U01