
Midmark M11 - v1.0.5
Total Cycles: 3611

Name: RITTER2
Sterilizer ID: V2313228
05/13/2024 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.2 F 0.0
2:00 88.5 F 0.0
4:00 107.9 F 0.1
6:00 143.0 F 1.2
8:00 163.6 F 3.7
10:00 177.9 F 6.7
12:00 193.6 F 9.8
14:00 205.7 F 13.1
16:00 216.6 F 6.2
18:00 224.2 F 10.0
20:00 233.5 F 10.5
22:00 237.5 F 11.0
24:00 234.8 F 8.7
26:00 247.2 F 14.5
28:00 257.3 F 20.0
30:00 266.7 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.4 F 30.0
2:00 272.5 F 29.6
3:00 271.3 F 29.1
4:00 271.3 F 29.5

Min 271.1 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 214.4 F 0.0
10:00 192.0 F 0.0
15:00 190.4 F 0.0
20:00 188.9 F 0.0
25:00 186.8 F 0.0
30:00 186.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

05/13/2024 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____