
Midmark M11 - v1.0.0
Total Cycles: 1800

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2022 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.7 F 0.0
2:00 117.1 F 0.0
4:00 141.6 F 2.1
6:00 166.4 F 5.6
8:00 197.8 F 9.6
10:00 218.6 F 7.4
12:00 225.1 F 7.0
14:00 235.8 F 11.5
16:00 232.7 F 7.7
18:00 245.2 F 13.4
20:00 243.0 F 11.7
22:00 255.5 F 18.2
24:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.7 F 29.0
2:00 271.3 F 28.2
3:00 272.7 F 29.1
4:00 272.0 F 28.4

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 227.1 F 0.0
10:00 217.7 F 0.0
15:00 214.8 F 0.0
20:00 221.2 F 0.0
25:00 235.3 F 0.0
30:00 239.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 25:34
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:30
CYCLE COMPLETE

04/06/2022 02:16 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____