
Midmark M11 - v1.0.5
Total Cycles: 2956

Name: RITTER2
Sterilizer ID: V2313228
10/02/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 179.5 F 3.6
4:00 198.1 F 9.1
6:00 216.8 F 7.6
8:00 229.8 F 10.2
10:00 232.6 F 7.9
12:00 241.9 F 11.6
14:00 245.7 F 13.5
16:00 258.4 F 20.3
18:00 269.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.4
2:00 271.5 F 28.9
3:00 272.6 F 29.5
4:00 271.6 F 28.7

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 212.3 F 0.0
10:00 203.0 F 0.0
15:00 202.5 F 0.0
20:00 200.2 F 0.0
25:00 216.1 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

10/02/2023 11:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____