
Midmark M11 - v1.0.0
Total Cycles: 4098

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2023 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.5 F 0.0
2:00 139.7 F 1.8
4:00 148.2 F 4.8
6:00 175.9 F 8.2
8:00 202.8 F 12.6
10:00 215.8 F 4.9
12:00 223.7 F 9.3
14:00 236.3 F 10.6
16:00 238.0 F 10.5
18:00 235.8 F 8.3
20:00 248.7 F 14.6
22:00 260.0 F 21.0
24:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.5 F 28.4
3:00 272.0 F 29.1
4:00 272.4 F 29.0

Min 271.2 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 234.9 F 0.0
10:00 229.7 F 0.0
15:00 222.4 F 0.0
20:00 225.4 F 0.0
25:00 226.8 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:50
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

12/08/2023 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____