
Midmark M11 - v1.0.5
Total Cycles: 3254

Name: RITTER2
Sterilizer ID: V2313228
12/27/2023 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 171.6 F 1.9
4:00 171.1 F 5.5
6:00 192.7 F 10.4
8:00 212.0 F 4.6
10:00 226.4 F 10.2
12:00 231.0 F 7.7
14:00 243.5 F 13.4
16:00 243.8 F 12.5
18:00 257.0 F 19.6
20:00 268.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.4
2:00 271.4 F 29.0
3:00 272.6 F 29.5
4:00 271.7 F 28.8

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 242.4 F 0.0
10:00 236.7 F 0.0
15:00 231.4 F 0.0
20:00 227.1 F 0.0
25:00 229.8 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

12/27/2023 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____