
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001526
11:44 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:34
STERILIZING 3:40
136.4 C 226kPa 3:40
136.2 C 224kPa 4:10
136.3 C 225kPa 4:40
136.4 C 226kPa 5:10
136.3 C 225kPa 5:40
136.4 C 226kPa 6:10
136.3 C 226kPa 6:40
136.4 C 227kPa 7:10
136.4 C 227kPa 7:10
VENTING 7:10
AIR DRYING 8:05
CYCLE COMPLETE 8:12

Digital Signature #
1475B4F6E1BD994A
