
Midmark M11 - v1.0.0
Total Cycles: 1200

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2021 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 183.5 F 3.6
4:00 203.6 F 8.2
6:00 222.0 F 10.3
8:00 233.5 F 9.0
10:00 238.7 F 10.2
12:00 244.3 F 12.5
14:00 242.7 F 11.4
16:00 256.2 F 18.4
18:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.6
2:00 271.4 F 28.2
3:00 272.3 F 28.7
4:00 271.4 F 28.0

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 206.0 F 0.0
10:00 195.8 F 0.0
15:00 195.4 F 0.0
20:00 192.7 F 0.0
25:00 189.8 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 2:44
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

09/23/2021 12:08 PM
ACCEPTED: X
REJECTED:
Operator: U02