
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 003707
10:51 27/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:54
STERILIZING 2:41
136.0 C 220kPa 2:41
136.4 C 225kPa 3:11
136.2 C 220kPa 3:41
136.3 C 221kPa 4:11
136.4 C 225kPa 4:41
136.1 C 221kPa 5:11
136.3 C 225kPa 5:41
136.1 C 223kPa 6:11
136.1 C 223kPa 6:11
VENTING 6:11
AIR DRYING 6:59
CYCLE COMPLETE 7:04

Digital Signature #
8B2C9280792A7913
