

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002289  
9:42 25/08/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:42  
PRESSURIZING 1:46  
STERILIZING 2:48  
136.5 C 225kPa 2:48  
136.2 C 223kPa 3:18  
136.1 C 218kPa 3:48  
136.4 C 222kPa 4:18  
136.4 C 223kPa 4:48  
136.4 C 224kPa 5:18  
136.4 C 224kPa 5:48  
136.3 C 225kPa 6:18  
136.3 C 225kPa 6:18  
VENTING 6:18  
AIR DRYING 7:12  
CYCLE COMPLETE 8:44  
-----

Digital Signature #  
56A5121A850D2987  
-----