
Midmark M11 - v1.0.0
Total Cycles: 3026

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.6 F 0.0
2:00 198.3 F 4.3
4:00 215.6 F 9.7
6:00 227.5 F 6.2
8:00 236.7 F 11.2
10:00 236.4 F 8.9
12:00 239.7 F 9.3
14:00 249.9 F 15.0
16:00 262.6 F 22.3
18:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.8
2:00 271.5 F 28.1
3:00 272.0 F 28.9
4:00 272.2 F 28.7

Min 271.2 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 233.3 F 0.0
10:00 231.0 F 0.0
15:00 228.4 F 0.0
20:00 228.1 F 0.0
25:00 229.7 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:09
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

03/22/2023 01:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____