
Midmark M11 - v1.0.0
Total Cycles: 766

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2021 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.7 F 0.0
2:00 121.3 F 0.3
4:00 130.2 F 2.0
6:00 161.7 F 4.8
8:00 194.1 F 8.3
10:00 210.1 F 13.1
12:00 219.0 F 6.8
14:00 230.2 F 11.6
16:00 232.4 F 8.2
18:00 242.4 F 11.6
20:00 244.2 F 12.5
22:00 256.2 F 18.7
24:00 267.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.3
2:00 272.0 F 28.7
3:00 271.1 F 28.5
4:00 272.3 F 29.3

Min 271.1 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 205.6 F 0.0
10:00 212.2 F 0.0
15:00 206.8 F 0.0
20:00 216.3 F 0.0
25:00 229.2 F 0.0
30:00 232.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:54
CYCLE COMPLETE

04/01/2021 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____