

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1504

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2022 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.9 F 0.0  
2:00 76.3 F 0.0  
4:00 83.6 F 0.0  
6:00 193.2 F 0.1  
8:00 155.3 F 2.5  
10:00 162.4 F 6.1  
12:00 200.9 F 10.1  
14:00 226.7 F 7.2  
16:00 222.6 F 7.1  
18:00 231.4 F 11.1  
20:00 232.5 F 7.3  
22:00 243.5 F 13.0  
24:00 242.0 F 11.5  
26:00 254.0 F 17.7  
28:00 264.8 F 24.1  
30:00 270.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.1 F 29.2  
2:00 272.2 F 29.2  
3:00 272.7 F 29.3  
4:00 272.8 F 29.6

Min 271.6 F 29.1  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 236.7 F 0.0  
10:00 226.5 F 0.0  
15:00 217.9 F 0.0  
20:00 212.3 F 0.0  
25:00 210.5 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 30:39  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:46  
CYCLE COMPLETE

09/20/2022 07:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_