
Midmark M11 - v1.0.5
Total Cycles: 4090

Name: RITTER2
Sterilizer ID: V2313228
11/05/2024 06:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 194.6 F 4.9
4:00 217.3 F 11.9
6:00 229.0 F 8.5
8:00 237.4 F 9.9
10:00 243.6 F 12.0
12:00 247.4 F 14.2
14:00 261.6 F 22.2
16:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.8
2:00 272.9 F 29.4
3:00 271.9 F 29.0
4:00 272.4 F 28.9

Min 271.6 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 210.0 F 0.0
10:00 225.7 F 0.0
15:00 219.4 F 0.0
20:00 214.2 F 0.0
25:00 211.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 3:07
HEATING: 16:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:24
CYCLE COMPLETE

11/05/2024 06:59 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____