
Midmark M11 - v1.0.5
Total Cycles: 94

Name: RITTER2
Sterilizer ID: V2313228
06/14/2021 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.6 F 0.0
2:00 169.6 F 2.6
4:00 173.9 F 6.3
6:00 194.3 F 10.9
8:00 213.9 F 5.3
10:00 224.0 F 10.7
12:00 230.5 F 8.0
14:00 240.4 F 11.9
16:00 244.0 F 13.2
18:00 256.0 F 19.5
20:00 266.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.6
1:00 272.0 F 30.1
2:00 271.3 F 29.7
3:00 272.0 F 30.1
4:00 271.4 F 29.4

Min 271.1 F 29.4
Max 272.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 229.7 F 0.0
10:00 230.7 F 0.0
15:00 223.0 F 0.0
20:00 217.6 F 0.0
25:00 213.2 F 0.0
Dry Cycle Aborted

FILLING: 1:24
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:15
DRYING: 26:43
TOTAL CYCLE: 00:56:57

06/14/2021 10:48 AM
ACCEPTED: X
REJECTED:
Operator: U01