
Midmark M11 - v1.0.0
Total Cycles: 3253

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 07:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.0 F 0.0
2:00 81.7 F 0.0
4:00 99.9 F 0.0
6:00 119.3 F 0.0
8:00 149.4 F 1.3
10:00 177.7 F 4.9
12:00 198.4 F 8.5
14:00 221.1 F 12.2
16:00 219.8 F 5.4
18:00 229.1 F 9.4
20:00 237.9 F 11.6
22:00 239.1 F 10.6
24:00 235.0 F 7.9
26:00 249.0 F 14.6
28:00 260.2 F 20.9
30:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.0 F 28.7
2:00 272.4 F 28.7
3:00 271.5 F 28.1
4:00 272.7 F 29.1

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 202.1 F 0.0
10:00 215.3 F 0.0
15:00 216.4 F 0.0
20:00 218.2 F 0.0
25:00 224.6 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

05/18/2023 08:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____