
Midmark M11 - v1.0.5
Total Cycles: 3649

Name: RITTER2
Sterilizer ID: V2313228
05/27/2024 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.1 F 0.0
2:00 129.5 F 1.5
4:00 148.8 F 4.4
6:00 173.8 F 8.0
8:00 195.1 F 12.2
10:00 211.6 F 5.4
12:00 227.1 F 10.0
14:00 230.1 F 8.8
16:00 237.9 F 11.7
18:00 237.1 F 10.0
20:00 249.6 F 16.1
22:00 260.5 F 22.3
24:00 270.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.7 F 30.4
2:00 272.0 F 29.8
3:00 271.2 F 29.7
4:00 272.6 F 30.5

Min 271.2 F 29.5
Max 272.9 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 211.9 F 0.0
10:00 206.8 F 0.0
15:00 199.8 F 0.0
20:00 199.2 F 0.0
25:00 202.0 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:52
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

05/27/2024 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____