
Midmark M11 - v1.0.0
Total Cycles: 940

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2021 04:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 174.2 F 2.1
4:00 175.7 F 4.9
6:00 195.5 F 8.7
8:00 209.0 F 13.1
10:00 224.8 F 7.3
12:00 235.6 F 12.0
14:00 234.8 F 9.0
16:00 241.2 F 10.5
18:00 245.9 F 13.3
20:00 257.4 F 19.4
22:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 272.1 F 28.6
3:00 271.5 F 28.6
4:00 272.8 F 29.2

Min 271.4 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 232.9 F 0.0
10:00 222.9 F 0.0
15:00 220.3 F 0.0
20:00 220.3 F 0.0
25:00 220.1 F 0.0
30:00 196.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

06/24/2021 05:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____