
Midmark M11 - v1.0.5
Total Cycles: 3105

Name: RITTER2
Sterilizer ID: V2313228
11/09/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.0 F 0.0
2:00 142.5 F 1.5
4:00 156.8 F 4.5
6:00 178.7 F 8.5
8:00 199.4 F 13.2
10:00 215.6 F 7.1
12:00 227.7 F 11.9
14:00 233.1 F 9.0
16:00 238.7 F 10.6
18:00 245.3 F 13.7
20:00 257.2 F 20.0
22:00 267.9 F 26.8
24:00 271.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.1 F 29.6
2:00 272.5 F 29.9
3:00 272.5 F 30.2
4:00 271.7 F 30.0

Min 271.6 F 29.6
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 206.5 F 0.0
10:00 210.7 F 0.0
15:00 210.1 F 0.0
20:00 205.2 F 0.0
25:00 201.8 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:03
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

11/09/2023 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____