
Midmark M11 - v1.0.5
Total Cycles: 4292

Name: RITTER2
Sterilizer ID: V2313228
01/24/2025 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 87.0 F 0.0
4:00 104.0 F 0.1
6:00 119.0 F 1.1
8:00 153.1 F 4.4
10:00 174.9 F 8.2
12:00 204.7 F 12.5
14:00 215.2 F 4.5
16:00 226.7 F 8.2
18:00 236.1 F 12.4
20:00 233.9 F 8.5
22:00 244.5 F 12.7
24:00 243.4 F 12.0
26:00 255.2 F 18.2
28:00 266.1 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.3 F 28.9
2:00 272.3 F 29.3
3:00 271.8 F 28.5
4:00 272.8 F 29.3

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 210.4 F 0.0
10:00 199.0 F 0.0
15:00 190.2 F 0.0
20:00 188.9 F 0.0
25:00 188.0 F 0.0
30:00 185.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:38
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:07:53
CYCLE COMPLETE

01/24/2025 10:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____