

-----  
-----  
Midmark M11 - v1.0.5  
Total Cycles: 90

Name: RITTER2  
Sterilizer ID: V2313228  
06/10/2021 02:17 PM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.5 F 0.0  
2:00 166.4 F 3.2  
4:00 181.9 F 7.3  
6:00 202.8 F 12.3  
8:00 215.1 F 6.7  
10:00 230.0 F 11.9  
12:00 233.8 F 9.6  
14:00 237.1 F 9.5  
16:00 248.2 F 15.2  
18:00 260.1 F 21.8  
20:00 271.1 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 271.7 F 29.3  
2:00 272.1 F 29.6  
3:00 272.1 F 29.9

Min 271.3 F 29.2  
Max 272.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.8 F 0.0  
5:00 208.3 F 0.0  
10:00 209.0 F 0.0  
15:00 208.8 F 0.0  
20:00 207.4 F 0.0  
25:00 208.1 F 0.0  
30:00 207.3 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 20:38  
STERILIZING: 3:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

06/10/2021 03:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_