
Midmark M11 - v1.0.5
Total Cycles: 4144

Name: RITTER2
Sterilizer ID: V2313228
11/20/2024 05:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 162.1 F 1.3
4:00 183.1 F 5.1
6:00 210.3 F 9.6
8:00 217.0 F 10.2
10:00 228.6 F 8.7
12:00 237.7 F 13.6
14:00 238.8 F 11.1
16:00 237.3 F 9.7
18:00 250.7 F 16.1
20:00 262.0 F 22.6
22:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.5
2:00 271.3 F 29.3
3:00 272.6 F 29.5
4:00 271.1 F 29.1

Min 271.0 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 238.9 F 0.0
10:00 225.5 F 0.0
15:00 204.9 F 0.0
20:00 180.7 F 0.0
25:00 177.2 F 0.0
30:00 175.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

11/20/2024 06:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____