
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001900
14:23 28/06/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CYCLE ABORTED 0:02
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:02

Digital Signature #
6B57690EE3641483
