
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000125
12:13 11/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:17
PRESSURIZING 4:17
STERILIZING 5:20
136.0 C 220kPa 5:20
136.4 C 219kPa 5:50
136.5 C 223kPa 6:20
136.4 C 220kPa 6:50
136.3 C 223kPa 7:20
136.4 C 224kPa 7:50
136.3 C 224kPa 8:20
136.4 C 226kPa 8:50
136.4 C 226kPa 8:50
VENTING 8:50
AIR DRYING 9:49
CYCLE COMPLETE 9:54

Digital Signature #
78C7AC144311AEF0
