
Midmark M11 - v1.0.0
Total Cycles: 5182

Name: RITTER 1
Sterilizer ID: V2178198
12/12/2024 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.9 F 0.0
2:00 187.3 F 0.3
4:00 201.6 F 2.4
6:00 204.4 F 7.0
8:00 221.0 F 11.5
10:00 230.1 F 6.5
12:00 237.8 F 11.8
14:00 237.5 F 9.2
16:00 238.3 F 9.6
18:00 247.8 F 13.9
20:00 259.3 F 20.2
22:00 269.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.6 F 28.5
2:00 271.8 F 28.5
3:00 272.5 F 28.7
4:00 272.6 F 28.9

Min 271.2 F 28.5
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 241.1 F 0.0
10:00 228.3 F 0.0
15:00 232.0 F 0.0
20:00 230.5 F 0.0
25:00 227.5 F 0.0
30:00 232.7 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

12/12/2024 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____