
Midmark M11 - v1.0.5
Total Cycles: 2260

Name: RITTER2
Sterilizer ID: V2313228
04/18/2023 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 186.5 F 3.1
4:00 199.1 F 8.1
6:00 219.1 F 12.1
8:00 228.2 F 9.6
10:00 233.5 F 9.0
12:00 243.7 F 13.0
14:00 244.0 F 12.9
16:00 256.9 F 19.7
18:00 268.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.6
2:00 271.3 F 29.2
3:00 272.3 F 29.7
4:00 271.4 F 29.0

Min 271.2 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 222.9 F 0.0
10:00 237.3 F 0.0
15:00 228.9 F 0.0
20:00 225.1 F 0.0
25:00 218.9 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

04/18/2023 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____