

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001825  
9:26 11/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 2:02  
STERILIZING 3:03  
136.0 C 220kPa 3:03  
136.3 C 224kPa 3:33  
136.3 C 224kPa 4:03  
136.3 C 225kPa 4:33  
136.4 C 222kPa 5:03  
136.4 C 225kPa 5:33  
136.4 C 225kPa 6:03  
136.0 C 223kPa 6:33  
136.0 C 223kPa 6:33  
VENTING 6:34  
AIR DRYING 7:23  
CYCLE COMPLETE 37:23  
-----

Digital Signature #  
CB13F6229F086813  
-----