

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2411

Name: RITTER2  
Sterilizer ID: V2313228  
05/24/2023 04:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
2:00 188.3 F 4.2  
4:00 202.1 F 9.5  
6:00 221.0 F 7.1  
8:00 231.7 F 10.7  
10:00 233.4 F 8.5  
12:00 239.7 F 10.2  
14:00 248.4 F 14.8  
16:00 261.2 F 22.0  
18:00 272.6 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.4 F 29.2  
2:00 272.2 F 29.5  
3:00 271.9 F 28.9  
4:00 272.4 F 29.6  
  
Min 271.5 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 225.8 F 0.0  
10:00 241.1 F 0.0  
15:00 232.5 F 0.0  
20:00 225.8 F 0.0  
25:00 224.3 F 0.0  
30:00 224.9 F 0.0  
  
Drying Complete

FILLING: 2:17  
HEATING: 18:22  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

05/24/2023 04:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_