
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002030
7:39 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:47
STERILIZING 3:49
136.2 C 222kPa 3:49
136.4 C 225kPa 4:19
136.3 C 224kPa 4:49
136.4 C 225kPa 5:19
136.4 C 221kPa 5:49
136.3 C 223kPa 6:19
136.4 C 224kPa 6:49
136.3 C 221kPa 7:19
136.3 C 221kPa 7:19
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 38:10

Digital Signature #
06F1ECBDB1D941F4
