
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008022
8:48 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 4:06
STERILIZING 5:15
136.4 C 220kPa 5:15
136.3 C 222kPa 5:45
136.1 C 221kPa 6:15
136.3 C 221kPa 6:45
136.4 C 223kPa 7:15
136.3 C 220kPa 7:45
136.1 C 221kPa 8:15
136.4 C 222kPa 8:45
136.4 C 222kPa 8:45
VENTING 8:45
AIR DRYING 9:37
CYCLE COMPLETE 39:37

Digital Signature #
F69E1DEF3467FF63
