
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001974
7:41 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:50
STERILIZING 3:53
136.0 C 220kPa 3:53
136.0 C 220kPa 4:23
136.2 C 221kPa 4:53
136.1 C 225kPa 5:23
136.4 C 223kPa 5:53
136.4 C 224kPa 6:23
136.4 C 225kPa 6:53
136.4 C 223kPa 7:23
136.4 C 223kPa 7:23
VENTING 7:23
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
1FF169C4419E7B9B
