
Midmark M11 - v1.0.0
Total Cycles: 3851

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.2 F 0.0
2:00 192.5 F 3.3
4:00 206.7 F 8.5
6:00 221.8 F 11.5
8:00 229.0 F 8.5
10:00 239.7 F 12.1
12:00 241.9 F 11.7
14:00 241.0 F 10.7
16:00 254.5 F 17.5
18:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.3 F 28.3
2:00 272.7 F 28.8
3:00 271.6 F 28.2
4:00 272.4 F 28.5

Min 271.6 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 233.5 F 0.0
10:00 214.4 F 0.0
15:00 207.5 F 0.0
20:00 202.0 F 0.0
25:00 197.9 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

10/02/2023 03:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____