

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001990  
7:11 12/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:21  
PRESSURIZING 1:27  
STERILIZING 2:30  
136.5 C 221kPa 2:30  
136.4 C 223kPa 3:00  
136.4 C 223kPa 3:30  
136.2 C 220kPa 4:00  
136.4 C 221kPa 4:30  
136.1 C 221kPa 5:00  
136.0 C 220kPa 5:30  
136.1 C 220kPa 6:00  
136.1 C 220kPa 6:00  
VENTING 6:01  
AIR DRYING 6:51  
CYCLE COMPLETE 6:54

-----  
Digital Signature #  
7F216781801DE30B  
-----