
Midmark M11 - v1.0.0
Total Cycles: 3964

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.5 F 0.0
2:00 172.1 F 2.2
4:00 171.2 F 5.7
6:00 194.4 F 10.1
8:00 221.6 F 5.4
10:00 228.0 F 9.1
12:00 239.9 F 10.5
14:00 240.4 F 11.0
16:00 236.7 F 8.9
18:00 250.3 F 15.4
20:00 261.6 F 21.9
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.1 F 29.0
2:00 271.7 F 28.4
3:00 271.7 F 28.8
4:00 272.8 F 29.2

Min 271.0 F 28.1
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 184.4 F 0.0
10:00 195.6 F 0.0
15:00 195.1 F 0.0
20:00 191.7 F 0.0
25:00 190.0 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

11/01/2023 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____