
Midmark M11 - v1.0.5
Total Cycles: 1309

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 72.6 F 0.0
4:00 79.2 F 0.0
6:00 101.9 F 0.0
8:00 138.3 F 2.0
10:00 165.6 F 5.2
12:00 185.8 F 8.6
14:00 202.6 F 12.3
16:00 218.1 F 5.4
18:00 228.1 F 9.3
20:00 237.3 F 12.5
22:00 237.1 F 10.5
24:00 235.5 F 8.6
26:00 246.7 F 14.2
28:00 257.6 F 20.1
30:00 267.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.1 F 29.9
2:00 272.4 F 29.7
3:00 271.7 F 29.1
4:00 271.3 F 29.2

Min 271.1 F 28.9
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 223.0 F 0.0
10:00 229.8 F 0.0
15:00 226.3 F 0.0
20:00 231.4 F 0.0
25:00 225.6 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:34
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

07/18/2022 08:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____