
Midmark M11 - v1.0.0
Total Cycles: 2349

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2022 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 178.3 F 3.9
4:00 190.0 F 8.1
6:00 208.7 F 13.1
8:00 219.5 F 7.1
10:00 233.2 F 12.3
12:00 234.7 F 9.2
14:00 237.1 F 9.2
16:00 247.2 F 14.0
18:00 259.6 F 20.9
20:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.5
2:00 272.2 F 29.0
3:00 271.2 F 28.6
4:00 272.0 F 29.3

Min 271.1 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 224.0 F 0.0
10:00 240.7 F 0.0
15:00 237.3 F 0.0
20:00 230.0 F 0.0
25:00 226.1 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

09/21/2022 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____