
Midmark M11 - v1.0.5
Total Cycles: 1600

Name: RITTER2
Sterilizer ID: V2313228
10/14/2022 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 152.6 F 2.8
4:00 176.3 F 6.3
6:00 196.0 F 11.1
8:00 215.0 F 5.1
10:00 225.7 F 10.5
12:00 229.3 F 7.4
14:00 242.2 F 13.2
16:00 241.7 F 11.9
18:00 254.1 F 18.3
20:00 265.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.1 F 30.0
2:00 272.1 F 29.5
3:00 271.5 F 29.6
4:00 272.3 F 29.9

Min 271.2 F 29.2
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 212.6 F 0.0
10:00 233.9 F 0.0
15:00 237.0 F 0.0
20:00 230.6 F 0.0
25:00 224.1 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

10/14/2022 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____