
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001430
12:14 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:48
STERILIZING 3:49
136.0 C 220kPa 3:49
136.4 C 225kPa 4:19
136.1 C 223kPa 4:49
136.3 C 224kPa 5:19
136.3 C 224kPa 5:49
136.4 C 224kPa 6:19
136.5 C 226kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:19
VENTING 7:19
AIR DRYING 8:13
CYCLE COMPLETE 8:17

Digital Signature #
0D842DDB9A2011A6
