

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3062

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2023 08:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.1 F 0.0  
2:00 83.2 F 0.0  
4:00 96.4 F 0.0  
6:00 129.6 F 0.1  
8:00 172.4 F 2.3  
10:00 191.4 F 6.4  
12:00 208.3 F 10.7  
14:00 226.9 F 5.5  
16:00 229.5 F 8.2  
18:00 239.0 F 12.4  
20:00 236.7 F 9.0  
22:00 240.6 F 9.7  
24:00 247.3 F 13.5  
26:00 259.6 F 20.2  
28:00 270.8 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 271.5 F 27.9  
2:00 271.5 F 27.8  
3:00 271.5 F 27.8  
4:00 271.7 F 27.7

Min 271.5 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.0 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 28:38  
STERILIZING: 4:00  
VENTING: 2:48  
DRYING: 0:19  
TOTAL CYCLE: 00:36:44

03/31/2023 08:53 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_