

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001238  
9:04 23/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:34  
STERILIZING 4:26  
136.0 C 220kPa 4:26  
135.8 C 219kPa 4:56  
136.1 C 222kPa 5:26  
136.1 C 222kPa 5:56  
136.1 C 221kPa 6:26  
136.2 C 222kPa 6:56  
136.1 C 221kPa 7:26  
136.3 C 224kPa 7:56  
136.3 C 224kPa 7:56  
VENTING 7:56  
AIR DRYING 8:51  
CYCLE COMPLETE 8:56

-----  
Digital Signature #  
BDAD43609E9126F9  
-----