
Midmark M11 - v1.0.0
Total Cycles: 3815

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 06:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 188.8 F 2.9
4:00 207.4 F 8.3
6:00 225.8 F 10.9
8:00 228.9 F 9.6
10:00 234.6 F 8.6
12:00 243.5 F 13.2
14:00 243.5 F 12.1
16:00 255.9 F 18.6
18:00 267.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 28.8
2:00 272.4 F 29.2
3:00 271.4 F 28.4
4:00 272.5 F 29.0

Min 271.1 F 28.3
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 201.7 F 0.0
10:00 201.6 F 0.0
15:00 195.9 F 0.0
20:00 192.2 F 0.0
25:00 189.6 F 0.0
30:00 186.7 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

09/25/2023 07:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____