
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001588
8:30 19/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 5:04
136.4 C 224kPa 5:04
136.4 C 224kPa 5:34
136.2 C 222kPa 6:04
135.9 C 219kPa 6:34
136.2 C 223kPa 7:04
136.3 C 223kPa 7:34
136.2 C 221kPa 8:04
136.3 C 224kPa 8:34
136.3 C 224kPa 8:34
VENTING 8:34
CYCLE ABORTED 9:10
STOP BUTTON PRESSED
NOT STERILE
END TIME 9:10

Digital Signature #
C66F595A61E3001E
