

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3213

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2023 01:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
2:00 173.2 F 3.7  
4:00 190.8 F 8.0  
6:00 222.9 F 13.0  
8:00 228.8 F 8.2  
10:00 239.7 F 13.5  
12:00 239.3 F 10.9  
14:00 236.7 F 9.0  
16:00 250.8 F 15.8  
18:00 262.6 F 22.5  
20:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.2 F 28.8  
2:00 272.2 F 29.2  
3:00 271.7 F 28.5  
4:00 272.5 F 29.2

Min 271.3 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 234.5 F 0.0  
10:00 224.6 F 0.0  
15:00 225.9 F 0.0  
20:00 213.2 F 0.0  
25:00 204.2 F 0.0  
30:00 201.0 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 20:16  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

05/10/2023 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_