
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000233
14:07 09/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:01
STERILIZING 3:59
136.4 C 224kPa 3:59
136.2 C 224kPa 4:29
136.1 C 222kPa 4:59
136.3 C 224kPa 5:29
136.4 C 224kPa 5:59
136.2 C 219kPa 6:29
136.4 C 225kPa 6:59
136.4 C 224kPa 7:29
136.4 C 224kPa 7:29
VENTING 7:30
AIR DRYING 8:20
CYCLE COMPLETE 9:39

Digital Signature #
D709E5F1EE48A746
