
Midmark M11 - v1.0.5
Total Cycles: 2876

Name: RITTER2
Sterilizer ID: V2313228
09/13/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.2 F 0.0
2:00 143.6 F 0.0
4:00 156.8 F 0.8
6:00 182.6 F 4.8
8:00 205.0 F 9.4
10:00 225.0 F 9.3
12:00 230.4 F 8.8
14:00 240.3 F 12.2
16:00 239.9 F 11.2
18:00 237.8 F 9.7
20:00 251.4 F 16.3
22:00 262.7 F 22.9
24:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.8
1:00 272.3 F 29.4
2:00 272.3 F 29.0
3:00 271.8 F 29.0
4:00 272.8 F 29.4

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 239.0 F 0.0
10:00 208.6 F 0.0
15:00 191.3 F 0.0
20:00 199.7 F 0.0
25:00 207.2 F 0.0
30:00 208.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

09/13/2023 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____