
Midmark M11 - v1.0.0
Total Cycles: 3051

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.2 F 0.0
2:00 186.3 F 2.6
4:00 191.7 F 7.2
6:00 211.4 F 12.3
8:00 225.1 F 7.3
10:00 238.8 F 12.9
12:00 238.4 F 10.5
14:00 236.3 F 8.8
16:00 251.4 F 16.1
18:00 263.8 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.6 F 28.3
3:00 272.5 F 29.2
4:00 272.4 F 28.9

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 217.4 F 0.0
10:00 237.6 F 0.0
15:00 234.5 F 0.0
20:00 230.4 F 0.0
25:00 222.7 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:60
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

03/29/2023 09:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____