
Midmark M11 - v1.0.5
Total Cycles: 3273

Name: RITTER2
Sterilizer ID: V2313228
01/03/2024 04:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 171.0 F 2.6
4:00 194.0 F 6.9
6:00 211.3 F 11.3
8:00 223.2 F 6.3
10:00 236.1 F 11.6
12:00 235.1 F 9.5
14:00 238.0 F 9.7
16:00 248.9 F 15.2
18:00 261.1 F 22.0
20:00 271.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.4 F 29.6
2:00 272.5 F 29.8
3:00 272.1 F 30.0
4:00 271.7 F 29.3

Min 271.0 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 228.5 F 0.0
10:00 214.2 F 0.0
15:00 187.5 F 0.0
20:00 183.8 F 0.0
25:00 181.3 F 0.0
30:00 179.6 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

01/03/2024 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____