
Midmark M11 - v1.0.0
Total Cycles: 3959

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.0 F 0.0
2:00 113.3 F 1.1
4:00 132.8 F 3.5
6:00 176.0 F 6.9
8:00 196.0 F 11.0
10:00 214.5 F 4.2
12:00 219.9 F 6.6
14:00 230.5 F 10.5
16:00 232.2 F 7.9
18:00 241.3 F 11.3
20:00 237.2 F 9.2
22:00 249.9 F 15.2
24:00 260.7 F 21.2
26:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 28.7
2:00 272.7 F 29.1
3:00 271.9 F 29.0
4:00 272.1 F 28.7

Min 271.4 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 250.1 F 0.0
10:00 225.2 F 0.0
15:00 228.0 F 0.0
20:00 219.8 F 0.0
25:00 233.5 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 27:23
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:06:30
CYCLE COMPLETE

10/30/2023 04:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____