
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001439
13:36 09/16/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:09
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:09

Digital Signature #
54D993BDA5B28F39
