
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 004459
15:21 27/09/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 4:40
STERILIZING 5:46
136.4 C 221kPa 5:46
136.2 C 222kPa 6:16
136.3 C 224kPa 6:46
136.4 C 224kPa 7:16
136.3 C 224kPa 7:46
136.4 C 223kPa 8:16
136.2 C 220kPa 8:46
136.4 C 225kPa 9:16
136.4 C 224kPa 9:46
136.4 C 225kPa 10:16
136.3 C 224kPa 10:46
136.4 C 224kPa 11:16
136.4 C 224kPa 11:46
136.3 C 224kPa 12:16
136.2 C 222kPa 12:46
136.4 C 225kPa 13:16
136.3 C 225kPa 13:46
136.4 C 225kPa 14:16
136.4 C 224kPa 14:46
136.4 C 225kPa 15:16
136.4 C 224kPa 15:46
136.4 C 224kPa 15:46
VENTING 15:46
CYCLE ABORTED 16:46
STOP BUTTON PRESSED
NOT STERILE

END TIME 16:46

Digital Signature #

358965FE9D904A93
